

# Общий белок (биреагент)

Биуретовый



## Параметры для ввода в программу анализатора Beckman-Coulter AU 680

Parameters		Specific Test Parameters			
General	LIH	ISE	HbA1c	Calculated Test	Range
Test Name: <input type="text" value="TP"/>		Type: <input type="text" value="Serum"/>		Operation: <input type="text" value="Yes"/>	
Sample Volume	<input type="text" value="2.4"/> $\mu\text{L}$	Dilution	<input type="text" value="0"/> $\mu\text{L}$	OD Limit	
Pre-Dilution Rate	<input type="text" value="1"/>			Min.OD	<input type="text" value="-2.0"/>
Rgt. Volume	R1(R1-1) <input type="text" value="120"/> $\mu\text{L}$	Dilution	<input type="text" value="0"/> $\mu\text{L}$	Max.OD	<input type="text" value="2.5"/>
	R2(R2-1) <input type="text" value="30"/> $\mu\text{L}$	Dilution	<input type="text" value="0"/> $\mu\text{L}$	Reagent OD Limit	
				First Low	<input type="text" value="-2.0"/>
				High	<input type="text" value="2.5"/>
				Last Low	<input type="text" value="-2.0"/>
				High	<input type="text" value="2.5"/>
Wavelength	Pri <input type="text" value="540"/> nm	Sec.	<input type="text" value="700"/> nm	Dynamic Range Low	<input type="text" value="0.5"/>
Method	<input type="text" value="END"/>			High	<input type="text" value="150"/>
Reaction Slope	<input type="text" value="+"/> $\nabla$			Correlation Factor A	<input type="text" value="1"/>
Measuring Point1 First	<input type="text" value="0"/>	Last	<input type="text" value="27"/>	Factor for Maker A	<input type="text" value="1"/>
Measuring Point2 First	<input type="text" value="0"/>	Last	<input type="text" value="10"/>	B	<input type="text" value="0"/>
Linearity Limit	<input type="text" value=""/>			B	<input type="text" value="0"/>
Lag Time Check	<input type="text" value=""/>			Onboard Stability Period	<input type="text" value="999"/> Day <input type="text" value=""/> Hour

Parameters		Specific Test Parameters			
General	LIH	ISE	HbA1c	Calculated Test	Range
Test Name: <input type="text" value="TP"/>		Type: <input type="text" value="Serum"/>			
Value/Flag:	<input type="text" value=""/>	Level L:	<input type="text" value=""/>	Level H:	<input type="text" value=""/>
Specific Ranges:					
	From	To		Low	High
<input type="checkbox"/> 1.	Sex <input type="text" value=""/>	Year <input type="text" value=""/>	Month <input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>
<input type="checkbox"/> 2.	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>
<input type="checkbox"/> 3.	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>
<input type="checkbox"/> 4.	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>
<input type="checkbox"/> 5.	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>
<input type="checkbox"/> 6.	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>
	7. No demographics			<input type="text" value="66"/>	<input type="text" value="88"/>
	8. Not within expected values			<input type="text" value="66"/>	<input type="text" value="88"/>
Unit	<input type="text" value="g/L"/>		Decimal Places	<input type="text" value="1"/>	
Panic Value					
Low <input type="text" value=""/> High <input type="text" value=""/>					

Parameters		Calibration Parameters			
Calibrators	Calibration Specific	STAT Table Calibration			
General	ISE				
Test Name: <input type="text" value="TP"/>		Type: <input type="text" value="Serum"/>		Use Serum Cal. <input type="checkbox"/>	
Calibration Type:	<input type="text" value="AB"/>	Formula:	<input type="text" value="Y=AX+B"/>	Counts:	<input type="text" value="2"/>
<Calibrator Parameters>					
Calibrator	OD	Conc	Factor/OD range	Low	High
Point 1:	<input type="text" value="TruCal U"/>	<input type="text" value="*"/>	<input type="text" value="-99999"/>	<input type="text" value="99999"/>	
Point 2:	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>	
Point 3:	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>	
Point 4:	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>	
Point 5:	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>	
Point 6:	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>	
Point 7:	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>	
Point 8:	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>	
Point 9:	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>	
Point 10:	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>	
Slope Check: <input type="text" value="None"/>					
Allowance Range Check					
<input type="checkbox"/> Reagent Blank <input type="text" value=""/>					
<input type="checkbox"/> Calibration <input type="text" value=""/>					
Advanced Calibration Operation: <input type="text" value=""/>					
Interval (RB/ACAL): <input type="text" value=""/>					
<Point Cal. For Master Curve>					
No. of Correction Points	<input type="text" value=""/>	Use Master Curve	<input type="checkbox"/>	<input type="checkbox"/> Lot Calibration	
Calibrator	OD	Conc	Low	High	Stability
Point 1:	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>	Reagent Blank <input type="text" value=""/> Day <input type="text" value=""/> Hour
Point 2:	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>	Calibration <input type="text" value=""/> Day <input type="text" value=""/> Hour
MB Type Factor:	<input type="text" value=""/>	1-Point Calibration Point	<input type="checkbox"/>	<input type="checkbox"/> with Conc-0	

Диапазон нормальных значений указан в соответствии с рекомендациями производителя реагентов.

\*-вводится из паспорта к калибратору.

Калибровка линейная, АВ (калибратор TruCal U)

Контроль TruLab N и TruLab P.

# Общий белок (монореагент)

Биуретовый



## Параметры для ввода в программу анализатора Beckman-Coulter AU 680

Parameters		Specific Test Parameters			
General	LIH	ISE	HbA1c	Calculated Test	Range
Test Name: <input type="text" value="TP"/>		Type: <input type="text" value="Serum"/>		Operation: <input type="text" value="Yes"/>	
Sample Volume	<input type="text" value="3"/> $\mu\text{L}$	Dilution	<input type="text" value="0"/> $\mu\text{L}$	OD Limit	
Pre-Dilution Rate	<input type="text" value="1"/>			Min.OD	<input type="text" value="-2.0"/>
Rgt. Volume	R1(R1-1) <input type="text" value="150"/> $\mu\text{L}$	Dilution	<input type="text" value="0"/> $\mu\text{L}$	Max.OD	<input type="text" value="2.5"/>
				Reagent OD Limit	
				First Low	<input type="text" value="-2.0"/>
				High	<input type="text" value="2.5"/>
				Last Low	<input type="text" value="-2.0"/>
				High	<input type="text" value="2.5"/>
R2(R2-1)	<input type="text" value="0"/> $\mu\text{L}$	Dilution	<input type="text" value="0"/> $\mu\text{L}$	Dynamic Range Low	<input type="text" value="0.5"/>
				High	<input type="text" value="150"/>
Wavelength	Pri <input type="text" value="540"/> nm	Sec.	<input type="text" value="700"/> nm	Correlation Factor A	<input type="text" value="1"/>
Method	<input type="text" value="END"/>			Factor for Maker A	<input type="text" value="1"/>
Reaction Slope	<input type="text" value="+"/>			B	<input type="text" value="0"/>
Measuring Point1 First	<input type="text" value="0"/>	Last	<input type="text" value="27"/>	B	<input type="text" value="0"/>
Measuring Point2 First		Last			
Linearity Limit				Onboard Stability Period	<input type="text" value="999"/> Day <input type="text" value=""/> Hour
Lag Time Check					

  

Parameters		Specific Test Parameters			
General	LIH	ISE	HbA1c	Calculated Test	Range
Test Name: <input type="text" value="TP"/>		Type: <input type="text" value="Serum"/>			
Value/Flag:	<input type="text"/>	Level L:	<input type="text"/>	Level H:	<input type="text"/>
Specific Ranges:					
	From	To		Low	High
<input type="checkbox"/> 1.	Sex <input type="text"/>	Year <input type="text"/>	Month <input type="text"/>	Year <input type="text"/>	Month <input type="text"/>
<input type="checkbox"/> 2.	<input type="text"/>	<input type="text"/>	<input type="text"/>	<input type="text"/>	<input type="text"/>
<input type="checkbox"/> 3.	<input type="text"/>	<input type="text"/>	<input type="text"/>	<input type="text"/>	<input type="text"/>
<input type="checkbox"/> 4.	<input type="text"/>	<input type="text"/>	<input type="text"/>	<input type="text"/>	<input type="text"/>
<input type="checkbox"/> 5.	<input type="text"/>	<input type="text"/>	<input type="text"/>	<input type="text"/>	<input type="text"/>
<input type="checkbox"/> 6.	<input type="text"/>	<input type="text"/>	<input type="text"/>	<input type="text"/>	<input type="text"/>
<input type="checkbox"/> 7.	No demographics			<input type="text" value="66"/>	<input type="text" value="88"/>
<input type="checkbox"/> 8.	Not within expected values			<input type="text" value="66"/>	<input type="text" value="88"/>
Unit	<input type="text" value="g/L"/>		Decimal Places	<input type="text" value="1"/>	
Panic Value					
Low <input type="text"/> High <input type="text"/>					

  

Parameters		Calibration Parameters		
Calibrators	Calibration Specific	STAT Table Calibration		
General	ISE			
Test Name: <input type="text" value="TP"/>		Type: <input type="text" value="Serum"/>		<input type="checkbox"/> Use Serum Cal.
Calibration Type:	<input type="text" value="AB"/>	Formula:	<input type="text" value="Y=AX+B"/>	
<Calibrator Parameters>		Counts:	<input type="text" value="2"/>	
Calibrator	OD	Conc	Factor/OD range	Slope Check
Point 1:	<input type="text" value="TruCal U"/>	<input type="text" value="*"/>	Low <input type="text" value="-99999"/> High <input type="text" value="99999"/>	<input type="text" value="None"/>
Point 2:	<input type="text"/>	<input type="text"/>		
Point 3:	<input type="text"/>	<input type="text"/>		
Point 4:	<input type="text"/>	<input type="text"/>		
Point 5:	<input type="text"/>	<input type="text"/>		
Point 6:	<input type="text"/>	<input type="text"/>		
Point 7:	<input type="text"/>	<input type="text"/>		
Point 8:	<input type="text"/>	<input type="text"/>		
Point 9:	<input type="text"/>	<input type="text"/>		
Point 10:	<input type="text"/>	<input type="text"/>		
<Point Cal. For Master Curve>	No. of Correction Points	<input type="text"/>	Use Master Curve	<input type="checkbox"/>
			<input type="checkbox"/> Lot Calibration	
Calibrator	OD	Conc	Low	High
Point 1:	<input type="text"/>	<input type="text"/>	<input type="text"/>	<input type="text"/>
Point 2:	<input type="text"/>	<input type="text"/>	<input type="text"/>	<input type="text"/>
MB Type Factor:	<input type="text"/>	1-Point Calibration Point	<input type="checkbox"/>	<input type="checkbox"/> with Conc-0
			Stability	Reagent Blank <input type="text"/> Day <input type="text"/> Hour
				Calibration <input type="text"/> Day <input type="text"/> Hour
				Advanced Calibration Operation <input type="text"/>
				Interval (RB/ACAL) <input type="text"/>
				<input type="checkbox"/> Reagent Blank <input type="text"/>
				<input type="checkbox"/> Calibration <input type="text"/>

Диапазон нормальных значений указан в соответствии с рекомендациями производителя реагентов.

\*-вводится из паспорта к калибратору.

Калибровка линейная, АВ (калибратор TruCal U)

Контроль TruLab N и TruLab P.