

## Параметры для ввода в программу анализатора Beckman-Coulter AU 480

Parameters		Specific Test Parameters			
General	LIH	ISE	HbA1c	Calculated Test	Range
Test Name: <input type="text" value="CO3"/>		Type: <input type="text" value="Serum"/>		Operation: <input type="text" value="Yes"/>	
Sample Volume	<input type="text" value="1"/>	μL	Dilution	<input type="text" value="0"/>	μL
Pre-Dilution Rate	<input type="text" value="1"/>		OD Limit	<input type="text" value="-2.0"/>	<input type="text" value="2.5"/>
Rgt. Volume	R1(R1-1)	<input type="text" value="100"/>	μL	Dilution	<input type="text" value="0"/>
	R2(R2-1)	<input type="text" value="0"/>	μL	Dilution	<input type="text" value="0"/>
Wavelength	Pri	<input type="text" value="405"/>	nm	Sec.	<input type="text" value="600"/>
Method		<input type="text" value="FIXED"/>		Factor for Maker	A
Reaction Slope		<input type="text" value="+"/>		Dynamic Range Low	<input type="text" value="4"/>
Measuring Point1 First		<input type="text" value="5"/>		High	<input type="text" value="50"/>
Measuring Point2 First				Correlation Factor A	<input type="text" value="1"/>
Linearity Limit				Factor for Maker	A
Lag Time Check				B	<input type="text" value="0"/>
				Onboard Stability Period	<input type="text" value="999"/>
				Day	<input type="text" value=""/>
				Hour	<input type="text" value=""/>

  

Parameters		Specific Test Parameters			
General	LIH	ISE	HbA1c	Calculated Test	Range
Test Name: <input type="text" value="CO3"/>		Type: <input type="text" value="Serum"/>			
Value/Flag:	<input type="text" value=""/>	Level L:	<input type="text" value=""/>	Level H:	<input type="text" value=""/>
Specific Ranges:					
	From	To		Low	High
	Sex	Year	Month	Year	Month
<input type="checkbox"/>	1.	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>
<input type="checkbox"/>	2.	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>
<input type="checkbox"/>	3.	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>
<input type="checkbox"/>	4.	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>
<input type="checkbox"/>	5.	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>
<input type="checkbox"/>	6.	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>
	7. No demographics			<input type="text" value="22"/>	<input type="text" value="29"/>
	8. Not within expected values			<input type="text" value="22"/>	<input type="text" value="29"/>
Unit	<input type="text" value="mmol/L"/>	Decimal Places	<input type="text" value="1"/>	Panic Value	
				Low	High
				<input type="text" value=""/>	<input type="text" value=""/>

  

Parameters		Calibration Parameters		
Calibrators	Calibration Specific	STAT Table Calibration		
General	ISE			
Test Name: <input type="text" value="CO3"/>		Type: <input type="text" value="Serum"/>		<input type="checkbox"/> Use Serum Cal.
Calibration Type:	<input type="text" value="AB"/>	Formula:	<input type="text" value="Y=AX+B"/>	Counts:
<Calibrator Parameters>		Factor/OD range		Slope Check
Calibrator	OD	Conc	Low	High
Point 1:	<input type="text" value="Cal bcarb"/>	<input type="text" value="*"/>	<input type="text" value="-99999"/>	<input type="text" value="99999"/>
Point 2:	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>
Point 3:	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>
Point 4:	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>
Point 5:	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>
Point 6:	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>
Point 7:	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>
Point 8:	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>
Point 9:	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>
Point 10:	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>
<Point Cal. For	No. of Correction Points	<input type="text" value=""/>	Use Master Curve	<input type="checkbox"/>
Master Curve>			OD Range	<input type="checkbox"/> Lot Calibration
Calibrator	OD	Conc	Low	High
Point 1:	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>
Point 2:	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>
MB Type Factor:	<input type="text" value=""/>	1-Point Calibration Point	<input type="checkbox"/>	<input type="checkbox"/> with Conc-0
			Stability	Reagent Blank
			Day	<input type="text" value=""/>
			Hour	<input type="text" value=""/>
			Calibration	Day
			Day	<input type="text" value=""/>
			Hour	<input type="text" value=""/>

Диапазон нормальных значений указан в соответствии с рекомендациями производителя реагентов. При использовании единиц измерения, отличающихся от приведенных, убедитесь, что значения стандартов, контрольных материалов, диапазонов нормальных значений и линейности метода введены в этих же единицах.

\*-вводится из паспорта к калибратору (Bicarbonate Standard).  
Контроль по TruLab Bicarbonat (один уровень).