

Параметры для ввода в программу анализатора Beckman-Coulter AU 680

Parameters		Specific Test Parameters			
General	LIH	ISE	HbA1c	Calculated Test	Range
Test Name: <input type="text" value="Gluc"/>		Type: <input type="text" value="Serum"/>		Operation: <input type="text" value="Yes"/>	
Sample Volume	<input type="text" value="1.6"/>	μL	Dilution	<input type="text" value="0"/>	μL
Pre-Dilution Rate	<input type="text" value="1"/>		OD Limit	<input type="text" value="-2.0"/>	<input type="text" value="2.5"/>
Rgt. Volume	R1(R1-1)	<input type="text" value="160"/>	μL	Dilution	<input type="text" value="0"/>
	R2(R2-1)	<input type="text" value="0"/>	μL	Dilution	<input type="text" value="0"/>
Wavelength	Pri	<input type="text" value="520"/>	nm	Sec.	<input type="text" value="700"/>
Method		<input type="text" value="END"/>			
Reaction Slope		<input type="text" value="+"/>			
Measuring Point1	First	<input type="text" value="0"/>	Last	<input type="text" value="27"/>	
Measuring Point2	First	<input type="text"/>	Last	<input type="text"/>	
Linearity Limit		<input type="text"/>	%		
Lag Time Check		<input type="text"/>			
			Dynamic Range Low	<input type="text" value="0"/>	High
			Correlation Factor A	<input type="text" value="1"/>	B
			Factor for Maker A	<input type="text" value="1"/>	B
			Onboard Stability Period	<input type="text" value="999"/>	Day
				<input type="text"/>	Hour

Parameters		Specific Test Parameters			
General	LIH	ISE	HbA1c	Calculated Test	Range
Test Name: <input type="text" value="Gluc"/>		Type: <input type="text" value="Serum"/>			
Value/Flag:	<input type="text"/>	Level L:	<input type="text"/>	Level H:	<input type="text"/>
Specific Ranges:					
	Sex	Year	Month	Year	Month
<input type="checkbox"/>	1.	<input type="text"/>	<input type="text"/>	<input type="text"/>	<input type="text"/>
<input type="checkbox"/>	2.	<input type="text"/>	<input type="text"/>	<input type="text"/>	<input type="text"/>
<input type="checkbox"/>	3.	<input type="text"/>	<input type="text"/>	<input type="text"/>	<input type="text"/>
<input type="checkbox"/>	4.	<input type="text"/>	<input type="text"/>	<input type="text"/>	<input type="text"/>
<input type="checkbox"/>	5.	<input type="text"/>	<input type="text"/>	<input type="text"/>	<input type="text"/>
<input type="checkbox"/>	6.	<input type="text"/>	<input type="text"/>	<input type="text"/>	<input type="text"/>
<input type="checkbox"/>	7. No demographics			<input type="text" value="3.9"/>	<input type="text" value="6.4"/>
<input type="checkbox"/>	8. Not within expected values			<input type="text" value="3.9"/>	<input type="text" value="6.4"/>
Unit	<input type="text" value="mmol/L"/>		Decimal Places	<input type="text" value="2"/>	
Panic Value					
		Low	<input type="text"/>	High	<input type="text"/>

Parameters		Calibration Parameters			
Calibrators	Calibration Specific	STAT Table Calibration			
General	ISE				
Test Name: <input type="text" value="Gluc"/>		Type: <input type="text" value="Serum"/>		Use Serum Cal. <input type="checkbox"/>	
Calibration Type:	<input type="text" value="AB"/>	Formula:	<input type="text" value="Y=AX+B"/>		Counts: <input type="text" value="2"/>
<Calibrator Parameters>					
Calibrator	OD	Conc	Factor/OD range	Low	High
Point 1:	<input type="text" value="TruCal U"/>	<input type="text" value="*"/>	<input type="text" value="-99999"/>	<input type="text" value="99999"/>	
Point 2:	<input type="text"/>	<input type="text"/>	<input type="text"/>	<input type="text"/>	
Point 3:	<input type="text"/>	<input type="text"/>	<input type="text"/>	<input type="text"/>	
Point 4:	<input type="text"/>	<input type="text"/>	<input type="text"/>	<input type="text"/>	
Point 5:	<input type="text"/>	<input type="text"/>	<input type="text"/>	<input type="text"/>	
Point 6:	<input type="text"/>	<input type="text"/>	<input type="text"/>	<input type="text"/>	
Point 7:	<input type="text"/>	<input type="text"/>	<input type="text"/>	<input type="text"/>	
Point 8:	<input type="text"/>	<input type="text"/>	<input type="text"/>	<input type="text"/>	
Point 9:	<input type="text"/>	<input type="text"/>	<input type="text"/>	<input type="text"/>	
Point 10:	<input type="text"/>	<input type="text"/>	<input type="text"/>	<input type="text"/>	
<Point Cal. For		No. of Correction Points	<input type="text"/>	Use Master Curve	<input type="checkbox"/>
Master Curve>		OD Range	Low	High	Stability
Point 1:	<input type="text"/>	<input type="text"/>	<input type="text"/>	<input type="text"/>	Reagent Blank
Point 2:	<input type="text"/>	<input type="text"/>	<input type="text"/>	<input type="text"/>	Day
MB Type Factor:		<input type="text"/>	1-Point Calibration Point	<input type="checkbox"/>	Day
		<input type="checkbox"/>	with Conc-0	<input type="checkbox"/>	Hour
Slope Check: <input type="text" value="None"/>					
Allowance Range Check					
o Reagent Blank <input type="text"/>					
o Calibration <input type="text"/>					
Advanced Calibration Operation: <input type="text"/>					
Interval (RB/ACAL): <input type="text"/>					

Диапазон нормальных значений указан в соответствии с рекомендациями производителя реагентов к набору и может быть изменен пользователем в соответствии с местными требованиями.

Адаптации для других типов образца запрашивайте дополнительно.
Калибровка линейная, АВ (калибратор TruCal U, или стандарт из набора).

* -вводится из паспорта к калибратору (стандарт)

Контроль TruLab N и TruLab P.