

Параметры для ввода в программу анализатора Beckman-Coulter AU 480

Parameters		Specific Test Parameters			
General	LIH	ISE	HbA1c	Calculated Test	Range
Test Name: <input type="text" value="mkALB"/>		Type: <input type="text" value="Urine"/>		Operation: <input type="text" value="Yes"/>	
Sample Volume	<input type="text" value="4.5"/> μL	Dilution	<input type="text" value="0"/> μL	OD Limit	
Pre-Dilution Rate	<input type="text" value="1"/>			Min.OD	<input type="text" value="-2.0"/>
Rgt. Volume	R1(R1-1) <input type="text" value="90"/> μL	Dilution	<input type="text" value="0"/> μL	Max.OD	<input type="text" value="2.5"/>
	R2(R2-1) <input type="text" value="18"/> μL	Dilution	<input type="text" value="0"/> μL	Reagent OD Limit	
				First Low	<input type="text" value="-2.0"/>
				High	<input type="text" value="2.5"/>
				Last Low	<input type="text" value="-2.0"/>
				High	<input type="text" value="2.5"/>
Wavelength	Pri <input type="text" value="410"/> nm	Sec.	<input type="text" value="None"/>	Dynamic Range Low	<input type="text" value="3"/>
Method	<input type="text" value="END"/>			High	<input type="text" value="350"/>
Reaction Slope	<input type="text" value="+"/>			Correlation Factor A	<input type="text" value="1"/>
Measuring Point1 First	<input type="text" value="0"/>	Last	<input type="text" value="27"/>	Factor for Maker A	<input type="text" value="1"/>
Measuring Point2 First	<input type="text" value="0"/>	Last	<input type="text" value="10"/>	B	<input type="text" value="0"/>
Linearity Limit	<input type="text" value=""/>			B	<input type="text" value="0"/>
Lag Time Check	<input type="text" value=""/>			Onboard Stability Period	<input type="text" value="999"/> Day <input type="text" value=""/> Hour

Parameters		Specific Test Parameters					
General	LIH	ISE	HbA1c	Calculated Test	Range		
Test Name: <input type="text" value="mkALB"/>		Type: <input type="text" value="Urine"/>					
Value/Flag:	<input type="text" value=""/>	Level L:	<input type="text" value=""/>	Level H:	<input type="text" value=""/>		
Specific Ranges:							
	From	To		Low	High		
<input type="checkbox"/> 1.	Sex <input type="text" value=""/>	Year <input type="text" value=""/>	Month <input type="text" value=""/>	Year <input type="text" value=""/>	Month <input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>
<input type="checkbox"/> 2.	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>
<input type="checkbox"/> 3.	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>
<input type="checkbox"/> 4.	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>
<input type="checkbox"/> 5.	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>
<input type="checkbox"/> 6.	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>
<input type="checkbox"/> 7.	No demographics			<input type="text" value="0"/>	<input type="text" value="30"/>		
<input type="checkbox"/> 8.	Not within expected values			<input type="text" value="0"/>	<input type="text" value="30"/>		
Unit	<input type="text" value="mg/L"/>		Decimal Places	<input type="text" value="1"/>			

Parameters		Calibration Parameters			
Calibrators	Calibration Specific	STAT Table Calibration			
General	ISE				
Test Name: <input type="text" value="mkALB"/>		Type: <input type="text" value="Urine"/>		<input type="checkbox"/> Use Serum Cal.	
Calibration Type:	<input type="text" value="6AB"/>	Formula:	<input type="text" value="Spline"/>	Counts:	<input type="text" value="2"/>
<Calibrator Parameters>					
Calibrator	OD	Conc	Factor/OD range Low	High	Slope Check
Point 1:	<input type="text" value="Saline"/>	<input type="text" value="0"/>	<input type="text" value="-0.1"/>	<input type="text" value="2.5"/>	<input type="text" value="None"/>
Point 2:	<input type="text" value="TruCal ALB U -1"/>	<input type="text" value="*"/>	<input type="text" value="-0.1"/>	<input type="text" value="2.5"/>	Allowance Range Check
Point 3:	<input type="text" value="TruCal ALB U -2"/>	<input type="text" value="*"/>	<input type="text" value="-0.1"/>	<input type="text" value="2.5"/>	<input type="checkbox"/> Reagent Blank <input type="text" value=""/>
Point 4:	<input type="text" value="TruCal ALB U -3"/>	<input type="text" value="*"/>	<input type="text" value="-0.1"/>	<input type="text" value="2.5"/>	<input type="checkbox"/> Calibration <input type="text" value=""/>
Point 5:	<input type="text" value="TruCal ALB U -4"/>	<input type="text" value="*"/>	<input type="text" value="-0.1"/>	<input type="text" value="2.5"/>	Advanced Calibration
Point 6:	<input type="text" value="TruCal ALB U -5"/>	<input type="text" value="*"/>	<input type="text" value="-0.1"/>	<input type="text" value="2.5"/>	Operation <input type="text" value=""/>
Point 7:	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>	Interval (RB/ACAL) <input type="text" value=""/>
Point 8:	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>	
Point 9:	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>	
Point 10:	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>	
<Point Cal. For		No. of Correction Points	<input type="text" value=""/>	Use Master Curve	<input type="checkbox"/>
Master Curve>		OD Range	<input type="text" value=""/>	<input type="checkbox"/> Lot Calibration	
Calibrator	OD	Conc	Low	High	Stability
Point 1:	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>	Reagent Blank <input type="text" value=""/>
Point 2:	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>	Day <input type="text" value=""/>
					Hour <input type="text" value=""/>
					Day <input type="text" value=""/>
					Hour <input type="text" value=""/>
MB Type Factor:	<input type="text" value=""/>	1-Point Calibration Point	<input type="checkbox"/>	<input type="checkbox"/> with Conc-0	

Диапазон нормальных значений указан в соответствии с рекомендациями производителя.

При использовании единиц измерения, отличающихся от приведенных, убедитесь, что значения стандартов, контрольных материалов, диапазонов нормальных значений и линейности метода введены в этих же единицах.

* -вводится из паспорта к калибратору (TruCal Alb U CSF). Первая точка - физраствор.

Контроль по TruLab Urine уровень 1 и уровень 2.