

Параметры для ввода в программу анализатора Beckman-Coulter AU 480

Parameters		Specific Test Parameters			
General	LIH	ISE	HbA1c	Calculated Test	Range
Test Name: <input type="text" value="TPur"/>		Type: <input type="text" value="Urine"/>		Operation <input type="text" value="Yes"/>	
Sample Volume	<input type="text" value="2"/> μL	Dilution	<input type="text" value="0"/> μL	OD Limit	
Pre-Dilution Rate	<input type="text" value="1"/>			Min.OD	<input type="text" value="-2.0"/> Max.OD <input type="text" value="2.5"/>
Rgt. Volume	R1(R1-1) <input type="text" value="100"/> μL	Dilution	<input type="text" value="0"/> μL	Reagent OD Limit	
				First Low	<input type="text" value="-2.0"/> High <input type="text" value="2.5"/>
				Last Low	<input type="text" value="-2.0"/> High <input type="text" value="2.5"/>
	R2(R2-1) <input type="text" value="0"/> μL	Dilution	<input type="text" value="0"/> μL	Dynamic Range Low <input type="text" value="10"/> High <input type="text" value="3000"/>	
Wavelength	Pri <input type="text" value="600"/> nm	Sec.	<input type="text" value="800"/> nm	Correlation Factor A <input type="text" value="1"/> B <input type="text" value="0"/>	
Method	<input type="text" value="END"/>			Factor for Maker A <input type="text" value="1"/> B <input type="text" value="0"/>	
Reaction Slope	<input type="text" value="+"/>	Onboard Stability Period		<input type="text" value="999"/> Day <input type="text" value=""/> Hour	
Measuring Point1 First	<input type="text" value="0"/>	Last	<input type="text" value="27"/>		
Measuring Point2 First	<input type="text" value=""/>	Last	<input type="text" value=""/>		
Linearity Limit	<input type="text" value=""/>				
Lag Time Check	<input type="text" value=""/>				

Parameters		Specific Test Parameters			
General	LIH	ISE	HbA1c	Calculated Test	Range
Test Name: <input type="text" value="TPur"/>		Type: <input type="text" value="Urine"/>			
Value/Flag: <input type="text" value=""/>		Level L: <input type="text" value=""/>		Level H: <input type="text" value=""/>	
Specific Ranges:					
	Sex	Year	Month	Year	Month
<input type="checkbox"/>	1.	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>
<input type="checkbox"/>	2.	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>
<input type="checkbox"/>	3.	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>
<input type="checkbox"/>	4.	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>
<input type="checkbox"/>	5.	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>
<input type="checkbox"/>	6.	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>
7. No demographics				<input type="text" value="0"/>	<input type="text" value="140"/>
8. Not within expected values				<input type="text" value=""/>	<input type="text" value=""/>
Unit <input type="text" value="mg/L"/>		Decimal Places <input type="text" value="0"/>			
Panic Value					
		Low		High	
		<input type="text" value=""/>		<input type="text" value=""/>	

Parameters		Calibration Parameters			
Calibrators	Calibration Specific	STAT Table Calibration			
General	ISE				
Test Name: <input type="text" value="TPur"/>		Type: <input type="text" value="Urine"/>		<input type="checkbox"/> Use Serum Cal.	
Calibration Type: <input type="text" value="AB"/>		Formula: <input type="text" value="Y=AX+B"/>		Counts: <input type="text" value="2"/>	
<Calibrator Parameters>					
Calibrator	OD	Conc	Factor/OD range		Slope Check
			Low	High	<input type="text" value="None"/>
Point 1:	<input type="text" value="StdTPur"/>	<input type="text" value="*"/>	<input type="text" value="-99999"/>	<input type="text" value="99999"/>	
Point 2:	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>	Allowance Range Check
Point 3:	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>	<input type="checkbox"/> Reagent Blank <input type="text" value=""/>
Point 4:	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>	<input type="checkbox"/> Calibration <input type="text" value=""/>
Point 5:	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>	Advanced Calibration
Point 6:	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>	Operation <input type="text" value=""/>
Point 7:	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>	Interval (RB/ACAL) <input type="text" value=""/>
Point 8:	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>	
Point 9:	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>	
Point 10:	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>	
<Point Cal. For		No. of Correction Points <input type="text" value=""/>	<input type="checkbox"/> Use Master Curve <input type="checkbox"/>		<input type="checkbox"/> Lot Calibration
<Master Curve>					
Calibrator	OD	Conc	Low	High	Stability
Point 1:	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>	Reagent Blank <input type="text" value=""/> Day <input type="text" value=""/> Hour
Point 2:	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>	Calibration <input type="text" value=""/> Day <input type="text" value=""/> Hour
MB Type Factor: <input type="text" value=""/>		1-Point Calibration Point <input type="text" value=""/>		<input type="checkbox"/> with Conc-0	

Диапазон нормальных значений указан в соответствии с рекомендациями производителя.

При использовании единиц измерения, отличающихся от приведенных, убедитесь, что значения стандартов, контрольных материалов, диапазонов нормальных значений и линейности метода введены в этих же единицах.

* вводится из инструкции к набору (или этикетки а флаконе стандарта)

Тип калибровки: линейная (Total protein UC standard (Стандарт общего белка в моче).

Контроль по TruLab Urine уровень 1 и уровень 2.