

Параметры для ввода в программу анализатора Beckman-Coulter AU 680

Parameters		Specific Test Parameters			
General	LIH	ISE	HbA1c	Calculated Test	Range
Test Name: CRP ▾		Type: Serum ▾		Operation Yes ▾	
Sample Volume	<input type="text" value="8"/> μL	Dilution	<input type="text" value="0"/> μL	OD Limit	
Pre-Dilution Rate	<input type="text" value="1"/> ▾			Min.OD	<input type="text" value="-2.0"/>
Rgt. Volume	R1(R1-1) <input type="text" value="125"/> μL	Dilution	<input type="text" value="0"/> μL	Max.OD	<input type="text" value="2.5"/>
				Reagent OD Limit	
				First Low	<input type="text" value="-2.0"/>
				High	<input type="text" value="2.5"/>
				Last Low	<input type="text" value="-2.0"/>
				High	<input type="text" value="2.5"/>
				Dynamic Range Low	<input type="text" value="0"/>
				High	<input type="text" value="350"/>
				Correlation Factor A	<input type="text" value="1"/>
				B	<input type="text" value="0"/>
				Factor for Maker A	<input type="text" value="1"/>
				B	<input type="text" value="0"/>
Wavelength	Pri <input type="text" value="340"/> nm ▾	Sec.	<input type="text" value="800"/> nm ▾	Onboard Stability Period	<input type="text" value="999"/> Day <input type="text" value=""/> Hour
Method	<input type="text" value="END"/> ▾				
Reaction Slope	<input type="text" value="+"/> ▾				
Measuring Point1 First	<input type="text" value="0"/>	Last	<input type="text" value="27"/>		
Measuring Point2 First	<input type="text" value="0"/>	Last	<input type="text" value="10"/>		
Linearity Limit	<input type="text" value=""/> %				
Lag Time Check	<input type="text" value=""/> ▾				

Parameters		Specific Test Parameters			
General	LIH	ISE	HbA1c	Calculated Test	Range
Test Name: CRP ▾		Type: Serum ▾			
Value/Flag:	<input type="text" value=""/> ▾	Level L:	<input type="text" value=""/>	Level H:	<input type="text" value=""/>
Specific Ranges:					
	From	To		Low	High
<input type="checkbox"/> 1.	Sex ▾	Year	Month	Year	Month
<input type="checkbox"/> 2.	<input type="text" value=""/> ▾	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>
<input type="checkbox"/> 3.	<input type="text" value=""/> ▾	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>
<input type="checkbox"/> 4.	<input type="text" value=""/> ▾	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>
<input type="checkbox"/> 5.	<input type="text" value=""/> ▾	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>
<input type="checkbox"/> 6.	<input type="text" value=""/> ▾	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>
<input type="checkbox"/> 7.	No demographics			<input type="text" value="0"/>	<input type="text" value="5"/>
<input type="checkbox"/> 8.	Not within expected values			<input type="text" value="0"/>	<input type="text" value="5"/>
Unit	<input type="text" value="mg/L"/>		Decimal Places	<input type="text" value="1"/>	
Panic Value					
Low <input type="text" value=""/> High <input type="text" value=""/>					

Parameters		Calibration Parameters		
Calibrators	Calibration Specific	STAT Table Calibration		
General	ISE			
Test Name: CRP ▾		Type: Serum ▾		Use Serum Cal. <input type="radio"/>
Calibration Type:	<input type="text" value="6AB"/> ▾	Formula:	<input type="text" value="Spline"/> ▾	Counts: <input type="text" value="2"/> ▾
<Calibrator Parameters>				
Calibrator	OD	Conc	Factor/OD range	Slope Check
			Low High	<input type="text" value="None"/> ▾
Point 1:	<input type="text" value="Saline"/>	<input type="text" value="0"/>	<input type="text" value="-0.1"/> <input type="text" value="2.5"/>	
Point 2:	<input type="text" value="TruCal CRP-1"/>	<input type="text" value="*"/>	<input type="text" value="-0.1"/> <input type="text" value="2.5"/>	Allowance Range Check
Point 3:	<input type="text" value="TruCal CRP-2"/>	<input type="text" value="*"/>	<input type="text" value="-0.1"/> <input type="text" value="2.5"/>	<input type="radio"/> Reagent Blank <input type="text" value=""/>
Point 4:	<input type="text" value="TruCal CRP-3"/>	<input type="text" value="*"/>	<input type="text" value="-0.1"/> <input type="text" value="2.5"/>	<input type="radio"/> Calibration <input type="text" value=""/>
Point 5:	<input type="text" value="TruCal CRP-4"/>	<input type="text" value="*"/>	<input type="text" value="-0.1"/> <input type="text" value="2.5"/>	Advanced Calibration
Point 6:	<input type="text" value="TruCal CRP-5"/>	<input type="text" value="*"/>	<input type="text" value="-0.1"/> <input type="text" value="2.5"/>	Operation <input type="text" value=""/> ▾
Point 7:	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>	Interval (RB/ACAL) <input type="text" value=""/> ▾
Point 8:	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>	
Point 9:	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>	
Point 10:	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>	
<Point Cal. For Master Curve>				
	No. of Correction Points	<input type="text" value=""/> ▾	Use Master Curve <input type="checkbox"/> ▾	<input type="radio"/> Lot Calibration
Calibrator	OD	Conc	Low High	Stability
Point 1:	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>	Reagent Blank <input type="text" value=""/> Day <input type="text" value=""/> Hour
Point 2:	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>	Calibration <input type="text" value=""/> Day <input type="text" value=""/> Hour
MB Type Factor:	<input type="text" value=""/> 1-Point Calibration Point	<input type="text" value=""/> ▾	<input type="radio"/> with Conc-0	

Диапазон нормальных значений указан в соответствии с рекомендациями производителя.

При использовании единиц измерения, отличающихся от приведенных, убедитесь, что значения стандартов, контрольных материалов, диапазонов нормальных значений и линейности метода введены в этих же единицах.

*-вводится в соответствии с настройками системы

** -вводится из паспорта к калибратору (TruCal CRP).

Контроль по TruLab Protein уровень 1 и уровень 2 или по TruLab CRP уровень 1 и 2.