

Параметры для ввода в программу анализатора Beckman-Coulter AU 680

Parameters		Specific Test Parameters			
General	LIH	ISE	HbA1c	Calculated Test	Range
Test Name: <input type="text" value="CO3"/>		Type: <input type="text" value="Serum"/>		Operation: <input type="text" value="Yes"/>	
Sample Volume	<input type="text" value="1.6"/>	μL	Dilution	<input type="text" value="0"/>	μL
Pre-Dilution Rate	<input type="text" value="1"/>		OD Limit	<input type="text" value="-2.0"/>	<input type="text" value="2.5"/>
Rgt. Volume	R1(R1-1)	<input type="text" value="160"/>	μL	Dilution	<input type="text" value="0"/>
	R2(R2-1)	<input type="text" value="0"/>	μL	Dilution	<input type="text" value="0"/>
Wavelength	Pri	<input type="text" value="405"/>	nm	Sec.	<input type="text" value="600"/>
Method	<input type="text" value="FIXED"/>				
Reaction Slope	<input type="text" value="+"/>		Onboard Stability Period: <input type="text" value="999"/> Day <input type="text" value=""/> Hour		
Measuring Point1	First	<input type="text" value="5"/>	Last	<input type="text" value="27"/>	
Measuring Point2	First	<input type="text" value=""/>	Last	<input type="text" value=""/>	
Linearity Limit	<input type="text" value=""/>		%		
Lag Time Check	<input type="text" value=""/>				

Parameters		Specific Test Parameters			
General	LIH	ISE	HbA1c	Calculated Test	Range
Test Name: <input type="text" value="CO3"/>		Type: <input type="text" value="Serum"/>			
Value/Flag:	<input type="text" value=""/>	Level L:	<input type="text" value=""/>	Level H:	<input type="text" value=""/>
Specific Ranges:					
	Sex	Year	Month	Year	Month
<input type="checkbox"/>	1.	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>
<input type="checkbox"/>	2.	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>
<input type="checkbox"/>	3.	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>
<input type="checkbox"/>	4.	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>
<input type="checkbox"/>	5.	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>
<input type="checkbox"/>	6.	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>
7. No demographics				<input type="text" value="22"/>	<input type="text" value="29"/>
8. Not within expected values				<input type="text" value="22"/>	<input type="text" value="29"/>
Unit	<input type="text" value="mmol/L"/>		Decimal Places	<input type="text" value="1"/>	
Panic Value					
		Low	<input type="text" value=""/>	High	<input type="text" value=""/>

Parameters		Calibration Parameters			
Calibrators	Calibration Specific	STAT Table Calibration			
General	ISE				
Test Name: <input type="text" value="CO3"/>		Type: <input type="text" value="Serum"/>		Use Serum Cal. <input type="checkbox"/>	
Calibration Type:	<input type="text" value="AB"/>		Formula:	<input type="text" value="Y=AX+B"/>	
<Calibrator Parameters>		Factor/OD range		Counts: <input type="text" value="2"/>	
Calibrator	OD	Conc	Low	High	Slope Check
Point 1:	<input type="text" value="Cal bcarb"/>	<input type="text" value="*"/>	<input type="text" value="-99999"/>	<input type="text" value="99999"/>	<input type="text" value="None"/>
Point 2:	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>
Point 3:	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>
Point 4:	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>
Point 5:	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>
Point 6:	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>
Point 7:	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>
Point 8:	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>
Point 9:	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>
Point 10:	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>
<Point Cal. For	No. of Correction Points	<input type="text" value=""/>	Use Master Curve	<input type="text" value=""/>	<input type="checkbox"/> Lot Calibration
Master Curve>		OD Range		Stability	
Calibrator	OD	Conc	Low	High	Reagent Blank
Point 1:	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>
Point 2:	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>
MB Type Factor:	<input type="text" value=""/>	1-Point Calibration Point	<input type="text" value=""/>	<input type="checkbox"/> with Conc-0	

Диапазон нормальных значений указан в соответствии с рекомендациями производителя реагентов. При использовании единиц измерения, отличающихся от приведенных, убедитесь, что значения стандартов, контрольных материалов, диапазонов нормальных значений и линейности метода введены в этих же единицах.

*-вводится из паспорта к калибратору (Bicarbonate Standard).
Контроль по TruLab Bicarbonat (один уровень).