

Параметры для ввода в программу анализатора Beckman-Coulter AU 680

Parameters		Specific Test Parameters			
General	LIH	ISE	HbA1c	Calculated Test	Range
Test Name: <input type="text" value="Mg"/>				Type: <input type="text" value="Serum"/>	Operation: <input type="text" value="Yes"/>
Sample Volume	<input type="text" value="1.6"/>	μL	Dilution	<input type="text" value="0"/>	μL
Pre-Dilution Rate	<input type="text" value="1"/>		OD Limit	<input type="text" value="-2.0"/>	<input type="text" value="2.5"/>
Rgt. Volume	R1(R1-1)	<input type="text" value="160"/>	μL	Dilution	<input type="text" value="0"/>
	R2(R2-1)	<input type="text" value="0"/>	μL	Dilution	<input type="text" value="0"/>
Wavelength	Pri	<input type="text" value="540"/>	nm	Sec.	<input type="text" value="660"/>
Method		<input type="text" value="END"/>		Factor for Maker	A: <input type="text" value="1"/> B: <input type="text" value="0"/>
Reaction Slope		<input type="text" value="+"/>		Onboard Stability Period	<input type="text" value="999"/> Day <input type="text" value=""/> Hour
Measuring Point1	First	<input type="text" value="0"/>	Last	<input type="text" value="27"/>	
Measuring Point2	First	<input type="text" value=""/>	Last	<input type="text" value=""/>	
Linearity Limit		<input type="text" value=""/>	%		
Lag Time Check		<input type="text" value=""/>			

Parameters		Specific Test Parameters			
General	LIH	ISE	HbA1c	Calculated Test	Range
Test Name: <input type="text" value="Mg"/>				Type: <input type="text" value="Serum"/>	
Value/Flag:	<input type="text" value=""/>	Level L:	<input type="text" value=""/>	Level H:	<input type="text" value=""/>
Specific Ranges:					
	From	To		Low	High
<input type="checkbox"/> 1.	Sex <input type="text" value=""/>	Year <input type="text" value=""/>	Month <input type="text" value=""/>	Year <input type="text" value=""/>	Month <input type="text" value=""/>
<input type="checkbox"/> 2.	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>
<input type="checkbox"/> 3.	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>
<input type="checkbox"/> 4.	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>
<input type="checkbox"/> 5.	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>
<input type="checkbox"/> 6.	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>
<input type="checkbox"/> 7.	No demographics			<input type="text" value="0.73"/>	<input type="text" value="1.06"/>
<input type="checkbox"/> 8.	Not within expected values			<input type="text" value=""/>	<input type="text" value=""/>
Unit	<input type="text" value="mmol/L"/>		Decimal Places	<input type="text" value="2"/>	
Panic Value					
		Low	<input type="text" value=""/>	High	<input type="text" value=""/>

Parameters		Calibration Parameters			
Calibrators	Calibration Specific	STAT Table Calibration			
General	ISE				
Test Name: <input type="text" value="Mg"/>				Type: <input type="text" value="Serum"/>	<input type="checkbox"/> Use Serum Cal.
Calibration Type:	<input type="text" value="AB"/>	Formula:	<input type="text" value="Y=AX+B"/>		Counts: <input type="text" value="2"/>
<Calibrator Parameters>					
Calibrator	OD	Conc	Factor/OD range	Low	High
Point 1:	<input type="text" value="TruCal U"/>	<input type="text" value="*"/>	<input type="text" value="-99999"/>	<input type="text" value="99999"/>	
Point 2:	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>	
Point 3:	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>	
Point 4:	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>	
Point 5:	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>	
Point 6:	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>	
Point 7:	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>	
Point 8:	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>	
Point 9:	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>	
Point 10:	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>	
Slope Check: <input type="text" value="None"/>					
Allowance Range Check					
<input type="checkbox"/> Reagent Blank <input type="text" value=""/>					
<input type="checkbox"/> Calibration <input type="text" value=""/>					
Advanced Calibration Operation: <input type="text" value=""/>					
Interval (RB/ACAL): <input type="text" value=""/>					
<Point Cal. For Master Curve>					
No. of Correction Points: <input type="text" value=""/>		Use Master Curve: <input type="checkbox"/>		<input type="checkbox"/> Lot Calibration	
Calibrator	OD	Conc	Low	High	Stability
Point 1:	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>	Reagent Blank: <input type="text" value=""/> Day <input type="text" value=""/> Hour
Point 2:	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>	Calibration: <input type="text" value=""/> Day <input type="text" value=""/> Hour
MB Type Factor: <input type="text" value=""/> 1-Point Calibration Point: <input type="checkbox"/> <input type="checkbox"/> with Conc-0					

Диапазон нормальных значений указан в соответствии с рекомендациями производителя реагентов к набору и может быть изменен пользователем в соответствии с местными требованиями.

Калибровка линейная, АВ (калибратор TruCal U, или стандарт из набора).

* -вводится из паспорта к калибратору (стандарт)

Контроль TruLab N и TruLab P.