

Параметры для ввода в программу анализатора Beckman-Coulter AU 680

Parameters		Specific Test Parameters			
General	LIH	ISE	HbA1c	Calculated Test	Range
Test Name: Bil-D ▾		Type: Serum ▾		Operation Yes ▾	
Sample Volume	<input type="text" value="12"/> μL	Dilution	<input type="text" value="0"/> μL	OD Limit	
Pre-Dilution Rate	<input type="text" value="1"/> ▾			Min.OD	<input type="text" value="-2.0"/>
Rgt. Volume	R1(R1-1) <input type="text" value="120"/> μL	Dilution	<input type="text" value="0"/> μL	Max.OD	<input type="text" value="2.5"/>
				Reagent OD Limit	
				First Low	<input type="text" value="-2.0"/>
				High	<input type="text" value="2.5"/>
				Last Low	<input type="text" value="-2.0"/>
				High	<input type="text" value="2.5"/>
R2(R2-1)	<input type="text" value="30"/> μL	Dilution	<input type="text" value="0"/> μL	Dynamic Range Low	<input type="text" value="0"/>
				High	<input type="text" value="170"/>
Wavelength	Pri <input type="text" value="540"/> nm ▾	Sec.	<input type="text" value="700"/> nm ▾	Correlation Factor A	<input type="text" value="1"/>
Method	END ▾			Factor for Maker A	<input type="text" value="1"/>
Reaction Slope	<input type="text" value="+"/> ▾			B	<input type="text" value="0"/>
Measuring Point1 First	<input type="text" value="0"/>	Last	<input type="text" value="27"/>	B	<input type="text" value="0"/>
Measuring Point2 First	<input type="text" value="0"/>	Last	<input type="text" value="10"/>		
Linearity Limit	<input type="text" value=""/> %			Onboard Stability Period	<input type="text" value="999"/> Day <input type="text" value=""/> Hour
Lag Time Check	<input type="text" value=""/> ▾				

Parameters		Specific Test Parameters			
General	LIH	ISE	HbA1c	Calculated Test	Range
Test Name: Bil-D ▾		Type: Serum ▾			
Value/Flag:	<input type="text" value=""/> ▾	Level L:	<input type="text" value=""/>	Level H:	<input type="text" value=""/>
Specific Ranges:					
	From	To		Low	High
<input type="checkbox"/> 1.	Sex ▾	Year	Month	Year	Month
<input type="checkbox"/> 2.	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>
<input type="checkbox"/> 3.	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>
<input type="checkbox"/> 4.	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>
<input type="checkbox"/> 5.	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>
<input type="checkbox"/> 6.	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>
<input type="checkbox"/> 7.	No demographics			<input type="text" value="0"/>	<input type="text" value="3.4"/>
<input type="checkbox"/> 8.	Not within expected values			<input type="text" value="0"/>	<input type="text" value="3.4"/>
Unit	mkmol/L		Decimal Places	<input type="text" value="2"/>	
Panic Value					
Low <input type="text" value=""/> High <input type="text" value=""/>					

Parameters		Calibration Parameters		
Calibrators	Calibration Specific	STAT Table Calibration		
General	ISE			
Test Name: Bil-D ▾		Type: Serum ▾		<input type="radio"/> Use Serum Cal.
Calibration Type:	AB ▾	Formula:	Y=AX+B ▾	
<Calibrator Parameters>		Counts:	<input type="text" value="2"/> ▾	
Calibrator	OD	Conc	Factor/OD range	Slope Check
Point 1:	TruCal U	*	Low -99999 High 99999	None ▾
Point 2:				Allowance Range Check
Point 3:				<input type="radio"/> Reagent Blank <input type="text" value=""/>
Point 4:				<input type="radio"/> Calibration <input type="text" value=""/>
Point 5:				Advanced Calibration
Point 6:				Operation <input type="text" value=""/> ▾
Point 7:				Interval (RB/ACAL) <input type="text" value=""/> ▾
Point 8:				
Point 9:				
Point 10:				
<Point Cal. For Master Curve>	No. of Correction Points <input type="text" value=""/> ▾	Use Master Curve <input type="text" value=""/> ▾	<input type="radio"/> Lot Calibration	
Calibrator	OD	Conc	Low	High
Point 1:				Stability
Point 2:				Reagent Blank <input type="text" value=""/> Day <input type="text" value=""/> Hour
				Calibration <input type="text" value=""/> Day <input type="text" value=""/> Hour
MB Type Factor:	<input type="text" value=""/>	1-Point Calibration Point <input type="text" value=""/> ▾	<input type="radio"/> with Conc-0	

Диапазон нормальных значений указан в соответствии с рекомендациями производителя. При использовании единиц измерения, отличающихся от приведенных, убедитесь, что значения стандартов, контрольных материалов, диапазонов нормальных значений и линейности метода введены в этих же единицах.

*-вводится из паспорта к калибратору (TruCalU).

Контроль по TruLab N и TruLab P.