

Параметры для ввода в программу анализатора Beckman-Coulter AU 680

Parameters		Specific Test Parameters			
General	LIH	ISE	HbA1c	Calculated Test	Range
Test Name: PhLip ▾		Type: Serum ▾		Operation Yes ▾	
Sample Volume	<input type="text" value="1,6"/> μL	Dilution	<input type="text" value="0"/> μL	OD Limit	
Pre-Dilution Rate	<input type="text" value="1"/> ▾			Min.OD	<input type="text" value="-2.0"/>
Rgt. Volume	R1(R1-1) <input type="text" value="180"/> μL	Dilution	<input type="text" value="0"/> μL	Max.OD	<input type="text" value="2.5"/>
				Reagent OD Limit	
				First Low	<input type="text" value="-2.0"/>
				High	<input type="text" value="2.5"/>
				Last Low	<input type="text" value="-2.0"/>
				High	<input type="text" value="2.5"/>
				Dynamic Range Low	<input type="text" value="0.09"/>
				High	<input type="text" value="13.3"/>
Wavelength	Pri <input type="text" value="570"/> nm ▾	Sec.	<input type="text" value="700"/> nm ▾	Correlation Factor A	<input type="text" value="1"/>
Method	<input type="text" value="END"/> ▾			Factor for Maker A	<input type="text" value="1"/>
Reaction Slope	<input type="text" value="+"/> ▾			B	<input type="text" value="0"/>
Measuring Point1 First	<input type="text" value="0"/>	Last	<input type="text" value="27"/>	B	<input type="text" value="0"/>
Measuring Point2 First	<input type="text" value="0"/>	Last	<input type="text" value="10"/>		
Linearity Limit	<input type="text" value=""/> %				
Lag Time Check	<input type="text" value=""/> ▾				
				Onboard Stability Period	<input type="text" value="999"/> Day <input type="text" value=""/> Hour

Parameters		Specific Test Parameters			
General	LIH	ISE	HbA1c	Calculated Test	Range
Test Name: PhLip ▾		Type: Serum ▾			
Value/Flag:	<input type="text" value=""/> ▾	Level L:	<input type="text" value=""/>	Level H:	<input type="text" value=""/>
Specific Ranges:					
	From	To		Low	High
<input type="checkbox"/> 1.	Sex ▾	Year	Month	Year	Month
<input type="checkbox"/> 2.	<input type="text" value=""/> ▾	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>
<input type="checkbox"/> 3.	<input type="text" value=""/> ▾	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>
<input type="checkbox"/> 4.	<input type="text" value=""/> ▾	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>
<input type="checkbox"/> 5.	<input type="text" value=""/> ▾	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>
<input type="checkbox"/> 6.	<input type="text" value=""/> ▾	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>
<input type="checkbox"/> 7.	No demographics			<input type="text" value="1.61"/>	<input type="text" value="3.55"/>
<input type="checkbox"/> 8.	Not within expected values			<input type="text" value=""/>	<input type="text" value=""/>
Unit	<input type="text" value="mmol/L"/>	Decimal Places	<input type="text" value="2"/>	Panic Value	
				Low	High

Parameters		Calibration Parameters			
Calibrators	Calibration Specific	STAT Table Calibration			
General	ISE				
Test Name: PhLip ▾		Type: Serum ▾		Use Serum Cal. <input type="radio"/>	
Calibration Type:	<input type="text" value="AB"/> ▾	Formula:	<input type="text" value="Y=AX+B"/> ▾	Counts:	<input type="text" value="2"/> ▾
<Calibrator Parameters>					
Calibrator	OD	Conc	Factor/OD range	Slope Check	<input type="text" value="None"/> ▾
Point 1:	<input type="text" value="Std PhLip"/>	<input type="text" value="*"/>	Low <input type="text" value="-99999"/> High <input type="text" value="99999"/>	Allowance Range Check	
Point 2:	<input type="text" value=""/>	<input type="text" value=""/>		<input type="radio"/> Reagent Blank	<input type="text" value=""/>
Point 3:	<input type="text" value=""/>	<input type="text" value=""/>		<input type="radio"/> Calibration	<input type="text" value=""/>
Point 4:	<input type="text" value=""/>	<input type="text" value=""/>		Advanced Calibration	
Point 5:	<input type="text" value=""/>	<input type="text" value=""/>		Operation	<input type="text" value=""/> ▾
Point 6:	<input type="text" value=""/>	<input type="text" value=""/>		Interval (RB/ACAL)	<input type="text" value=""/> ▾
Point 7:	<input type="text" value=""/>	<input type="text" value=""/>		<Point Cal. For No. of Correction Points <input type="text" value=""/> ▾ Use Master Curve <input type="checkbox"/> ▾ <input type="radio"/> Lot Calibration	
Point 8:	<input type="text" value=""/>	<input type="text" value=""/>		Master Curve>	
Point 9:	<input type="text" value=""/>	<input type="text" value=""/>		Calibrator	OD
Point 10:	<input type="text" value=""/>	<input type="text" value=""/>		Conc	Low
				High	Stability
				Reagent Blank	<input type="text" value=""/> Day <input type="text" value=""/> Hour
				Calibration	<input type="text" value=""/> Day <input type="text" value=""/> Hour
MB Type Factor:	<input type="text" value=""/>	1-Point Calibration Point <input type="checkbox"/> ▾	<input type="checkbox"/> ▾	<input type="radio"/> with Conc-0	

Диапазон нормальных значений указан в соответствии с рекомендациями производителя. При использовании единиц измерения, отличающихся от приведенных, убедитесь, что значения стандартов, контрольных материалов, диапазонов нормальных значений и линейности метода введены в этих же единицах.

*-вводится из паспорта к калибратору, калибратор—Phospholipids standard.

Контроль по TruLab N и TruLab P, или по TruLab L (липидный) уровень 1 и уровень 2.