Аполипопротеин B

ImmunoTD (Иммунотурбидиметрический)

Параметры для ввода в программу анализатора Beckman-Coulter AU 680

Parameters | Specific Test Parameters

General | LIH | ISE | HbA1c | Calculated Test | Range

Test Name: ApoB ▽ < > Type: Serum ▽ Operation Yes ▽

Sample Volume	1.6 µL	Dilution	0 µL	OD Limit			
Pre-Dilution Rate	1 ▽			Min.OD	-2.0	Max.OD	2.5
Rgt. Volume R1(R1-1)	160 µL	Dilution	0 µL	Reagent OD Limit			
				First Low	-2.0	High	2.5
				Last Low	-2.0	High	2.5
R2(R2-1)	32 µL	Dilution	0 µL				
				Dynamic Range Low	0	High	2.5
				Correlation Factor A	1	B	0
Wavelength Pri	340 ▽nm	Sec.	None ▽nm	Factor for Maker A	1	B	0
Method	END ▽						
Reaction Slope	+ ▽			Onboard Stability Period	999 Day		Hour
Measuring Point1 First	0	Last	27				
Measuring Point2 First	0	Last	10				
Linearity Limit	%						
Lag Time Check	▽						

Parameters | Specific Test Parameters

General | LIH | ISE | HbA1c | Calculated Test | Range

Test Name: ApoB ▽ < > Type: Serum ▽

Value/Flag: ▽ Level L: Level H:

Specific Ranges:

Panic Value: Low | High

	Sex	From Year	From Month	To Year	To Month	Low	High
☐ 1.	M ▽					0.8	1.55
☐ 2.	F ▽					0.75	1.5
☐ 3.	▽						
☐ 4.	▽						
☐ 5.	▽						
☐ 6.	▽						
7. No demographics						0.75	1.55
8. Not within expected values						0.75	1.55

Unit g/L Decimal Places 2

Parameters | Calibration Parameters

Calibrators | Calibration Specific | STAT Table Calibration

General | ISE

Test Name: ApoB ▽ < > Type Serum ▽ o Use Serum Cal.

Calibration Type: 6AB ▽ Formula: Spline ▽ Counts: 2 ▽

<Calibrator Parameters>

	Calibrator	OD	Conc	Factor/OD range Low	High
Point 1:	Saline		0	-0.1	2.5
Point 2:	ApoA1/B-1		*	-0.1	2.5
Point 3:	ApoA1/B-2		*	-0.1	2.5
Point 4:	ApoA1/B-3		*	-0.1	2.5
Point 5:	ApoA1/B-4		*	-0.1	2.5
Point 6:	ApoA1/B-5		*	-0.1	2.5
Point 7					
Point 8					
Point 9					
Point 10					

Slope Check None ▽

Allowance Range Check

o Reagent Blank

o Calibration

Advanced Calibration Operation ▽

Interval (RB/ACAL) ▽

<Point Cal. For Master Curve> No. of Correction Points ▽ Use Master Curve ▽ o Lot Calibration

	Calibrator	OD	Conc	OD Range Low	High	Stability
Point 1:						Reagent Blank Day Hour
Point 2:						Calibration Day Hour

MB Type Factor: 1-Point Calibration Point ▽ o with Conc-0

Диапазон нормальных значений указан в соответствии с рекомендациями производителя реагентов. При использовании единиц измерения, отличающихся от приведенных, убедитесь, что значения стандартов, контрольных материалов, диапазонов нормальных значений и линейности метода введены в этих же единицах.

*-вводится из паспорта к калибратору (TruCal Apo A1/B, 5 уровней).

Контроль по TruLab L уровень 1 и уровень 2 или TruLab N и P.