

Параметры для ввода в программу анализатора Beckman-Coulter AU 680

Parameters		Specific Test Parameters			
General	LIH	ISE	HbA1c	Calculated Test	Range
Test Name: <input type="text" value="C3c"/>		Type: <input type="text" value="Serum"/>		Operation <input type="text" value="Yes"/>	
Sample Volume	<input type="text" value="1.8"/>	μL	Dilution	<input type="text" value="0"/>	μL
Pre-Dilution Rate	<input type="text" value="1"/>		OD Limit	<input type="text" value="-2.0"/>	Max.OD <input type="text" value="2.5"/>
Rgt. Volume	R1(R1-1) <input type="text" value="125"/>	μL	Dilution	<input type="text" value="0"/>	μL
	R2(R2-1) <input type="text" value="25"/>	μL	Dilution	<input type="text" value="0"/>	μL
Wavelength	Pri <input type="text" value="340"/>	nm	Sec. <input type="text" value="800"/>	nm	
Method	<input type="text" value="END"/>		Dynamic Range Low	<input type="text" value="0"/>	High <input type="text" value="5"/>
Reaction Slope	<input type="text" value="+"/>		Correlation Factor A	<input type="text" value="1"/>	B <input type="text" value="0"/>
Measuring Point1 First	<input type="text" value="0"/>		Factor for Maker A	<input type="text" value="1"/>	B <input type="text" value="0"/>
Measuring Point2 First	<input type="text" value="0"/>		Onboard Stability Period	<input type="text" value="999"/>	Day <input type="text" value=""/> Hour
Linearity Limit	<input type="text" value=""/>	%	Last	<input type="text" value="27"/>	
Lag Time Check	<input type="text" value=""/>		Last	<input type="text" value="10"/>	

Parameters		Specific Test Parameters			
General	LIH	ISE	HbA1c	Calculated Test	Range
Test Name: <input type="text" value="C3c"/>		Type: <input type="text" value="Serum"/>			
Value/Flag:	<input type="text" value=""/>	Level L:	<input type="text" value=""/>	Level H:	<input type="text" value=""/>
Specific Ranges:					
	From	To		Low	High
<input type="checkbox"/> 1.	Sex <input type="text" value=""/>	Year <input type="text" value=""/>	Month <input type="text" value=""/>	Year <input type="text" value=""/>	Month <input type="text" value=""/>
<input type="checkbox"/> 2.	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>
<input type="checkbox"/> 3.	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>
<input type="checkbox"/> 4.	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>
<input type="checkbox"/> 5.	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>
<input type="checkbox"/> 6.	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>
<input type="checkbox"/> 7.	No demographics			<input type="text" value="0.9"/>	<input type="text" value="1.8"/>
<input type="checkbox"/> 8.	Not within expected values			<input type="text" value="0.9"/>	<input type="text" value="1.8"/>
Unit	<input type="text" value="g/L"/>		Decimal Places	<input type="text" value="2"/>	

Parameters		Calibration Parameters		
Calibrators	Calibration Specific	STAT Table Calibration		
General	ISE			
Test Name: <input type="text" value="C3c"/>		Type: <input type="text" value="Serum"/>		<input type="checkbox"/> Use Serum Cal.
Calibration Type:	<input type="text" value="6AB"/>	Formula:	<input type="text" value="Spline"/>	Counts: <input type="text" value="2"/>
<Calibrator Parameters>				
Calibrator	OD	Conc	Factor/OD range Low High	Slope Check <input type="text" value="None"/>
Point 1:	Saline	0	-0.1 2.5	
Point 2:	TruCal Prot-1	*	-0.1 2.5	Allowance Range Check
Point 3:	TruCal Prot-2	*	-0.1 2.5	<input type="checkbox"/> Reagent Blank <input type="text" value=""/>
Point 4:	TruCal Prot-3	*	-0.1 2.5	<input type="checkbox"/> Calibration <input type="text" value=""/>
Point 5:	TruCal Prot-4	*	-0.1 2.5	Advanced Calibration
Point 6:	TruCal Prot-5	*	-0.1 2.5	Operation <input type="text" value=""/>
Point 7:				Interval (RB/ACAL) <input type="text" value=""/>
Point 8:				
Point 9:				
Point 10:				
<Point Cal. For Master Curve>				
Calibrator	OD	Conc	Low High	Stability
Point 1:				Reagent Blank <input type="text" value=""/> Day <input type="text" value=""/> Hour
Point 2:				Calibration <input type="text" value=""/> Day <input type="text" value=""/> Hour
MB Type Factor:	<input type="text" value=""/>	1-Point Calibration Point <input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/> with Conc-0

Диапазон нормальных значений указан в соответствии с рекомендациями производителя.

При использовании единиц измерения, отличающихся от приведенных, убедитесь, что значения стандартов, контрольных материалов, диапазонов нормальных значений и линейности метода введены в этих же единицах.

*-вводится в соответствии с настройками системы

** -вводится из паспорта к калибратору (TruCal Protein).

Контроль по TruLab Protein уровень 1 и уровень 2.