

LDL, Липопротеины низкой плотности, ЛПНП

Прямой, селестивный



Параметры для ввода в программу анализатора Beckman-Coulter AU 680

Parameters		Specific Test Parameters			
General	LIH	ISE	HbA1c	Calculated Test	Range
Test Name: <input type="text" value="LDL"/>		Type: <input type="text" value="Serum"/>		Operation: <input type="text" value="Yes"/>	
Sample Volume	<input type="text" value="1.6"/>	μL	Dilution	<input type="text" value="0"/>	μL
Pre-Dilution Rate	<input type="text" value="1"/>		OD Limit	<input type="text" value="-2.0"/>	<input type="text" value="2.5"/>
Rgt. Volume	R1(R1-1)	<input type="text" value="140"/>	μL	Dilution	<input type="text" value="0"/>
	R2(R2-1)	<input type="text" value="35"/>	μL	Dilution	<input type="text" value="0"/>
Wavelength	Pri	<input type="text" value="600"/>	nm	Sec.	<input type="text" value="700"/>
Method		<input type="text" value="END"/>		Dynamic Range Low	<input type="text" value="0"/>
Reaction Slope		<input type="text" value="+"/>		High	<input type="text" value="10.3"/>
Measuring Point1 First		<input type="text" value="0"/>		Correlation Factor A	<input type="text" value="1"/>
Measuring Point2 First		<input type="text" value="0"/>		Factor for Maker A	<input type="text" value="1"/>
Linearity Limit		<input type="text" value=""/>		Onboard Stability Period	<input type="text" value="999"/>
Lag Time Check		<input type="text" value=""/>		Day	<input type="text" value=""/>
				Hour	<input type="text" value=""/>

Parameters		Specific Test Parameters			
General	LIH	ISE	HbA1c	Calculated Test	Range
Test Name: <input type="text" value="LDL"/>		Type: <input type="text" value="Serum"/>			
Value/Flag:	<input type="text" value=""/>	Level L:	<input type="text" value=""/>	Level H:	<input type="text" value=""/>
Specific Ranges:					
	Sex	Year	Month	Year	Month
<input type="checkbox"/>	1.	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>
<input type="checkbox"/>	2.	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>
<input type="checkbox"/>	3.	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>
<input type="checkbox"/>	4.	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>
<input type="checkbox"/>	5.	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>
<input type="checkbox"/>	6.	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>
	7. No demographics			<input type="text" value="0"/>	<input type="text" value="3.4"/>
	8. Not within expected values			<input type="text" value=""/>	<input type="text" value=""/>
Unit	<input type="text" value="mmol/L"/>		Decimal Places	<input type="text" value="2"/>	

Parameters		Calibration Parameters			
Calibrators	Calibration Specific	STAT Table Calibration			
General	ISE				
Test Name: <input type="text" value="LDL"/>		Type: <input type="text" value="Serum"/>		Use Serum Cal. <input type="radio"/>	
Calibration Type:	<input type="text" value="AB"/>	Formula:	<input type="text" value="Y=AX+B"/>		Counts: <input type="text" value="2"/>
<Calibrator Parameters>					
Calibrator	OD	Conc	Factor/OD range	Slope Check	<input type="text" value="None"/>
Point 1:	<input type="text" value="TruCal HDL/LDL"/>	<input type="text" value="*"/>	<input type="text" value="-99999"/>	<input type="text" value="99999"/>	
Point 2:	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>	
Point 3:	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>	
Point 4:	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>	
Point 5:	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>	
Point 6:	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>	
Point 7:	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>	
Point 8:	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>	
Point 9:	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>	
Point 10:	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>	
Allowance Range Check					
<input type="radio"/> Reagent Blank <input type="text" value=""/>					
<input type="radio"/> Calibration <input type="text" value=""/>					
Advanced Calibration Operation <input type="text" value=""/>					
Interval (RB/ACAL) <input type="text" value=""/>					
<Point Cal. For No. of Correction Points <input type="text" value=""/>					
Use Master Curve <input type="text" value=""/>					
<input type="radio"/> Lot Calibration					
Master Curve					
Calibrator	OD	Conc	Low	High	Stability
Point 1:	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>	Reagent Blank <input type="text" value=""/>
Point 2:	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>	<input type="text" value=""/>	Day <input type="text" value=""/>
Calibration <input type="text" value=""/>					
Day <input type="text" value=""/>					
Hour <input type="text" value=""/>					
MB Type Factor: <input type="text" value=""/>					
1-Point Calibration Point <input type="text" value=""/>					
<input type="radio"/> with Conc-0					

Диапазон нормальных значений указан в соответствии с рекомендациями производителя. При использовании единиц измерения, отличающихся от приведенных, убедитесь, что значения стандартов, контрольных материалов, диапазонов нормальных значений и линейности метода введены в этих же единицах.

* -вводится из паспорта к калибратору, калибратор – TruCal HDL/LDL
Контроль по TruLab L (липидный) уровень 1 и уровень 2.