

Параметры для ввода в программу анализатора Beckman-Coulter AU 480

Parameters		Specific Test Parameters			
General	LIH	ISE	HbA1c	Calculated Test	Range
Test Name: Prealb ▾		Type: Serum ▾		Operation Yes ▾	
Sample Volume	<input type="text" value="1.1"/> μL	Dilution	<input type="text" value="0"/> μL	OD Limit	
Pre-Dilution Rate	<input type="text" value="1"/> ▾			Min.OD	<input type="text" value="-2.0"/>
Rgt. Volume	R1(R1-1) <input type="text" value="90"/> μL	Dilution	<input type="text" value="0"/> μL	Max.OD	<input type="text" value="2.5"/>
				Reagent OD Limit	
				First Low	<input type="text" value="-2.0"/>
				High	<input type="text" value="2.5"/>
				Last Low	<input type="text" value="-2.0"/>
				High	<input type="text" value="2.5"/>
R2(R2-1)	<input type="text" value="18"/> μL	Dilution	<input type="text" value="0"/> μL	Dynamic Range Low	<input type="text" value="0"/>
				High	<input type="text" value="1.5"/>
Wavelength	Pri <input type="text" value="410"/> nm ▾	Sec.	None ▾	Correlation Factor A	<input type="text" value="1"/>
Method	END ▾			Factor for Maker A	<input type="text" value="1"/>
Reaction Slope	<input type="text" value="+"/> ▾			B	<input type="text" value="0"/>
Measuring Point1 First	<input type="text" value="0"/>	Last	<input type="text" value="27"/>	B	<input type="text" value="0"/>
Measuring Point2 First	<input type="text" value="0"/>	Last	<input type="text" value="10"/>		
Linearity Limit	<input type="text"/> %			Onboard Stability Period	<input type="text" value="999"/> Day <input type="text"/> Hour
Lag Time Check	<input type="text"/> ▾				

Parameters		Specific Test Parameters			
General	LIH	ISE	HbA1c	Calculated Test	Range
Test Name: Prealb ▾		Type: Serum ▾			
Value/Flag:	<input type="text"/> ▾	Level L:	<input type="text"/>	Level H:	<input type="text"/>
Specific Ranges:					
	From	To		Low	High
<input type="checkbox"/> 1.	Sex ▾	Year	Month	Year	Month
<input type="checkbox"/> 2.	<input type="text"/> ▾	<input type="text"/>	<input type="text"/>	<input type="text"/>	<input type="text"/>
<input type="checkbox"/> 3.	<input type="text"/> ▾	<input type="text"/>	<input type="text"/>	<input type="text"/>	<input type="text"/>
<input type="checkbox"/> 4.	<input type="text"/> ▾	<input type="text"/>	<input type="text"/>	<input type="text"/>	<input type="text"/>
<input type="checkbox"/> 5.	<input type="text"/> ▾	<input type="text"/>	<input type="text"/>	<input type="text"/>	<input type="text"/>
<input type="checkbox"/> 6.	<input type="text"/> ▾	<input type="text"/>	<input type="text"/>	<input type="text"/>	<input type="text"/>
<input type="checkbox"/> 7.	No demographics			<input type="text" value="0.2"/>	<input type="text" value="0.4"/>
<input type="checkbox"/> 8.	Not within expected values			<input type="text" value="0.2"/>	<input type="text" value="0.4"/>
Unit	<input type="text" value="g/L"/>	Decimal Places	<input type="text" value="2"/>	Panic Value	
				Low	High
				<input type="text"/>	<input type="text"/>

Parameters		Calibration Parameters		
Calibrators	Calibration Specific	STAT Table Calibration		
General	ISE			
Test Name: Prealb ▾		Type: Serum ▾		<input type="radio"/> Use Serum Cal.
Calibration Type:	6AB ▾	Formula:	Spline ▾	Counts: <input type="text" value="2"/> ▾
<Calibrator Parameters>				
Calibrator	OD	Conc	Factor/OD range	Slope Check
			Low High	None ▾
Point 1:	Saline	<input type="text" value="0"/>	-0.1 2.5	
Point 2:	TruCal Prot-1	*	-0.1 2.5	Allowance Range Check
Point 3:	TruCal Prot-2	*	-0.1 2.5	<input type="radio"/> Reagent Blank <input type="text"/>
Point 4:	TruCal Prot-3	*	-0.1 2.5	<input type="radio"/> Calibration <input type="text"/>
Point 5:	TruCal Prot-4	*	-0.1 2.5	Advanced Calibration
Point 6:	TruCal Prot-5	*	-0.1 2.5	Operation <input type="text"/> ▾
Point 7:				Interval (RB/ACAL) <input type="text"/> ▾
Point 8:				
Point 9:				
Point 10:				
<Point Cal. For Master Curve>				
	No. of Correction Points	<input type="text"/> ▾	Use Master Curve	<input type="radio"/> Lot Calibration
Calibrator	OD	Conc	Low High	Stability
Point 1:	<input type="text"/>	<input type="text"/>	<input type="text"/>	Reagent Blank <input type="text"/> Day <input type="text"/> Hour
Point 2:	<input type="text"/>	<input type="text"/>	<input type="text"/>	Calibration <input type="text"/> Day <input type="text"/> Hour
MB Type Factor:	<input type="text"/>	1-Point Calibration Point	<input type="text"/> ▾	<input type="radio"/> with Conc-0

Диапазон нормальных значений указан в соответствии с рекомендациями производителя.

При использовании единиц измерения, отличающихся от приведенных, убедитесь, что значения стандартов, контрольных материалов, диапазонов нормальных значений и линейности метода введены в этих же единицах.

* -вводится из паспорта к калибратору (TruCal Protein). Первая точка - физраствор.

Контроль по TruLab Protein уровень 1 и уровень 2.